

**NATIONAL INSTITUTE FOR HEALTH AND CARE
EXCELLENCE**

MEDICAL TECHNOLOGIES EVALUATION PROGRAMME

Equality impact assessment: Topic selection and scoping

Endocuff Vision for endoscopic investigation

The impact on equality has been assessed during this evaluation according to the principles of the NICE Equality scheme.

- | |
|---|
| 1. Have any potential equality issues been identified during the development of the briefing note or during the Committee meeting, and, if so, what are they? |
|---|

The EndoCuff device instructions for use state that it is not intended for deep ileal intubation and that it is contraindicated for people with acute, severe colitis or colonic strictures.
--

- | |
|---|
| 2. What is the preliminary view as to what extent these potential equality issues need addressing by the Committee? |
|---|

It is unlikely that these issues would hinder access of the technology for any specific group.
--

- | |
|--|
| 3. Does the scope need to highlight any potential equality issues? If so, how? |
|--|

No. The information above was included as a special consideration.
--

Approved by Acting Programme Director: Mark Campbell

Date: 10 May 2018